# Dexterity-enhanced Telerobotic Microsurgery

#### Steve Charles

MicroDexterity Systems, Inc. and Charles Retina Institute, Memphis, TN

Hari Das, Timothy Ohm, Curtis Boswell, Guillermo Rodriguez, Robert Steele Jet Propulsion Laboratory, California Institute of Technology, Pasadena, CA.

Dan Istrate
California Institute of Technology, Pasadena, CA.

Abstract - A telerobotic platform developed in a collaboration between NASA-JPL and MicroDexterity Systems, Inc (MDS) is described in this paper. The lightweight, compact 6 dof master-slave system is precise to better than 10 microns and can cover a workspace greater than 400 cubic centimeters. Current capabilities of the system include manual position control with augmented shared control modes and automatic modes of control of the robot. Simulated force feedback on the master device has been implemented and plans are to integrate force reflection from the slave end effector and evaluate the performance improvements enabled by the telerobot in simulated microsurgical tasks. The telerobot was used in a recent demonstration of a simulated eye microsurgical procedure.

### 1 Introduction

Microsurgery requires manipulation of tissue features of about fifty to a few hundred microns in size while visualizing the surgical field through a microscope. Surgeons manually manipulate instruments designed for specific parts of a procedure while viewing the instrument tips under the microscope under a 20 to 30 times magnification. While the tool for aiding the surgeon to see the microscopical field is already available, a similar tool for manipulating at this microscopic scale is currently absent. A microsurgical manipulator should scale down the surgeon's hand motions to the microscopic field and so allow the average surgeon to perform at the level of the best surgeons and allow the most skillful surgeons

to perform at unprecedented levels of dexterity [1].

This work is the result of a collaboration between researchers at the Jet Propulsion Laboratory and Steve Charles, MD, a vitreo-retinal surgeon. The Robot Assisted MicroSurgery (RAMS) telerobotic workstation developed at JPL [9] is a prototype of a system that will be completely under the manual control of a surgeon. The system, shown on Figure 1, has a slave robot that will hold surgical instruments. The slave robot




Figure 1: RAMS telerobot system.

motions replicate in six degrees of freedom those of the surgeon's hand measured using a master input device with a surgical instrument shaped handle. The surgeon commands motions for the instrument by moving the handle in the desired trajectories. The trajectories are

measured, filtered, and scaled down then used to drive the slave robot.

We present the details of this telerobotic system by first giving an overview of the subsystems and their interactions in the following section then present details in subsequent subsections divided according to subsystem. This paper concludes with a description of a recent demonstration of a simulated microsurgery procedure performed at JPL.

# 2 System description

Figure 2 shows an overview of the hardware components of the RAMS telerobotic system.




Figure 2: RAMS slave robot system.

Components of the RAMS system have been categorized into four subsystems. They are the mechanical subsystem, the electronics subsystem, the servo-control subsystem and the high-level software subsystem. The mechanical subsystem consists of a master input device and a slave robot arm with associated motors, encoders, gears, cables, pulleys and linkages that cause the tip of the robot to move under computer control and to measure the surgeon's hand motions precisely. The electronics subsystem consists of the motor amplifiers, a safety electronics circuit and relays within the amplifier box shown on Figure 2. These elements of the subsystem ensure that a number of error conditions are handled

gracefully. The servo-control subsystem is implemented in hardware and software. The relevant hardware parts of the subsystem are the servo-control boards and the computational processor boards. Servo-control software functions include setting-up the control parameters and running the servo-loop on the servo-control board to control the six motors, implementing the communication between the computation and servo-control boards, initializing the servo-control system and communicating with the electronics subsystem and communicating with the high-level software subsystem. The high-level software subsystem interfaces with a user, controls initialization of the system software and hardware, implements a number of demonstration modes of robot control and computes both the forward and inverse kinematics. A drawing of the interaction between the subsystems of the RAMS slave robot is shown on Figure 3.




Figure 3: Sub-systems of the RAMS telerobot system.

#### 2.1 Mechanical subsystem

The RAMS slave manipulator is a six degrees-of-freedom tendon-driven robotic arm designed to be compact yet exhibit very precise 10 micron relative positioning capability as well as maintain a very high work volume. Physically, the arm measures 2.5 cm. in diameter and is 25.0 cm. long from its base to tip. It is mounted to a cylindrical base housing which measures 12 cm. in diameter by 18 cm long that contains all of the drives that actuate the arm. A photograph of the arm appears on Figure 4. The joints of the arm are a torso joint rotating about an axis aligned with the base axis and positioned at the point the arm emerges from its base, a shoulder joint rotating about two axes that are in the same plane and perpendicular to the preceding

links, an elbow joint that also rotates about two axes that are in the same plane and perpendicular to the preceding links, and a wrist with pitch, yaw and roll joints.




Figure 4: RAMS slave robot.

The master device, kinematically similar to the slave robot, also has six tendon driven joints. It is 2.5 cm. in diameter and 25 cm. long. Its base houses high-resolution optical encoders requiring a larger volume a box of size 10.8 cm by 18.4 cm by 23.5 cm. Gear transformation ratios in the master arm are reduced to allow backdrivability. A photograph of the master input device is shown on Figure 5.

The slave wrist design (based on the kinematics of the Rosheim OMNI-WRIST [8]) utilizes a dual universal joint to give a three degrees-of-freedom, singularity free, mechanically decoupled joint that operates in a full hemisphere of motion (up to 90 degrees in any direction). The master wrist design uses a universal joint to transmit rotation motion through the joint while allowing pitch and yaw motions about the joint resulting in singularity free motion over a smaller range of motion in three degrees-of-freedom. The fourth and fifth axes of the master and slave robots are unique joints that rotate about 2 axes and allow passage of cables




Figure 5: RAMS Master input device.

to pass through the joint for actuating the succeeding joints without affecting their cable lengths. The sixth axes are torso joints which simply rotates the manipulators relative to their base housing, For the slave robot the torso range of motion is 330 degrees while on the master it is 30 degrees.

Features resulting from the unique mechanical design of the arms are:

- Drive Unit Separability Drive motors and optical encoders on the slave robot cannot survive an autoclave environment and are designed to be removable for sterilization.
- Zero/Low Backlash Low backlash (free play) is essential for doing fine manipulation, especially since the position sensors are on the motor shafts.
- Low Stiction Stiction (stick/slip characteristic) must be minimized to achieve small incremental movements without overshooting or instability.
- Decoupled Joints Having all joints mechanically decoupled simplifies kinematic computations as well as provides for partial functionality should one joint fail.
- Large Work Envelope A large work volume is desirable so that the slave arm's base will not have to be repositioned frequently during tasks.
- **High Stiffness** A stiff manipulator is necessary for accurate positioning under gravitational or en-

vironmental loads, especially when position sensing is at the motor drives.

- Backdrivability The master arm has been designed to be easily backdrivable.
- Compact/Lightweight In some applications, a restricted workspace warrants a small manipulator to minimize interference (i.e. visual interference).
- Fine Incremental Motions Human dexterity limitations constrain surgical procedures to feature sizes of about 20-50 microns, whereas the slave arm is designed to achieve better than 10 microns relative positioning accuracy. Notion of leveling the plane of dexterity\*\*\*

Ability to move has not matched view

- Precise position measurement The master arm has been designed to be able to measure commanded hand motions down to a relative position resolution of 25 microns, while the slave robot can read its tip position to a resolution of 1 micron.
- Tool Wiring Provisions Tools requiring electrical or pneumatic power can have it routed through a passageway through both the master and slave arms.

#### 2.2 Electronics subsystem

The RAMS electronics subsystem design includes off the shelf and custom designed electronics. Figure 6 shows a layout of its general components.




Figure 6: Electronics components and cabling.

Components of the electronics subsystem are a VME chassis, an amplifier chassis and safety electronics. The VME chassis houses the VME backplane and two Motorola MVME-167 computer boards used for high level system control. The VME chassis also contains the PMAC servo control cards and six supporting interface modules, power supplies (+/-15v) and a cable interface board. The VME chassis front panel contains main power control (AC) for the system. The rear panel provides access to the control computers serial communications port (RS-232). All components above are off the shelf items except the cable interface board.

The VME computer boards are the hardware portion of the high level control system. The RS-232 interface provides communication for control and observation of the robot system functions.

The PMAC servo boards generate 2 phase drive signals for sinusoidal commutation of the systems brushless DC motors. The PMAC receives optical encoder feedback from the motor shafts and provides low level control of the motors. The six I/O blocks and cable interface board handle signal and power distribution to the connectors on the rear panel.

The AMP (amplifier chassis) contains the six slave robot motor and three master robot motor drive amplifiers, system control electronics board, amplifier power supply and AMP subsystem power. The AMP chassis has interfaces to the VME chassis (analog inputs and control signals), the Slave robot (motor drive signals) and to the CTRL panel subsystem (panic stop, run and initialize). The AMP chassis main power (AC) is provided by the VME chassis.

The Amplifier sub-chassis secures the individual amplifiers to the AMP chassis. This is designed to provide a thermal path to the chassis and to provide a favorable orientation with respect to the chassis air flow pattern. The individual amplifiers should run cool. The frame of the Amplifier sub-chassis contains all necessary amplifier interface wiring. This makes the design highly modular to facilitate rapid check out and trouble-shooting.

The safety control electronics consists of the control electronics board and the brake relay board. The purpose of the braking function is to hold the motors in place when they are not under amplifier control. Programmable Logic Arrays (PLD) in the safety control electronics module monitors amplifier power, operator control buttons and the PANIC-HALT button, and a watchdog signal from the high-level software and control processors (indicating that they are healthy). Any anomaly triggers brakes to be set on the slave robot joint and a fault LED to be lighted. The operator must reset

the safety control electronics to re-activate the system. A diagram of the safety control electronics functions and PLD state transitions are shown on Figures 7 and 8.




Figure 7: Function of the control electronics.




Figure 8: Control electronics state transitions.

### 2.3 Servo-control subsystem

The RAMS servo-control system is implemented on processor boards and servo-control boards installed in a VME chassis. Two Motorola MVME-167 boards, named  $Proc\theta$  and Proc1, are installed on the VME chassis and run under the VxWorks operating system.  $Proc\theta$  performs kinematic, communication and high-level control functions. These functions are described in the High Level Software Architecture Section. Calls to subroutines that read and set joint angle positions of the robot are made from the high-level real-time software on  $Proc\theta$ . These routines, through shared memory implemented between  $Proc\theta$  and Proc1, provide setpoints

and read current joint angles of the robot. *Proc1*, in turn, passes the setpoints for controlling the robot to the servo control board and retrieves the joint angles measured by the servo-control board. The servo level control system uses the PMAC-VME board by Delta Tau.

Communication between *Proc0*, *Proc1* and the PMAC-VME boards is through shared memory. The PMAC board has a large variety of features for motor control, with a customer base largely from industrial installations. The key features used for control of the RAMS robot include:

- Digital sine-wave commutation.
- Automatic trajectory generation.
- Shared memory interface.
- Built-in amplifier/encoder interface.
- Robust closed loop control.

# 2.4 High-level software subsystem

There are a number of components to the high-level software for the RAMS slave robot. A drawing of the parts of the software in shown on Figure 9. Embedded in the




Figure 9: Parts of the high-level software.

computational blocks of the real-time control software are the kinematic control algorithms. They are based on algorithms developed at JPL [6],[7] for the unique geometry of the robot. The demonstration of different control modes of the robot was implemented using a software development tool for real-time systems called Control Shell [3],[4]. Handling of operator commands in the real-time software, transitions between states of control, changes in data flow due to transitions of states in the software and the algorithms executed within computation blocks. The user specifies the control modes of

the system through a graphic user interface (GUI) implemented with Tcl/Tk [2]. Commands entered into the GUI are transmitted over an Ethernet connection and are received on the real-time software side of the system. The message passing between the 2 parts of the software system uses NDDS [5]. A producer part creates the messages and broadcasts them from the GUI part of the system and a consumer part receives the messages and processes them.

# 3 Simulated Surgery

In September of 1996, a demonstration of a simulated eye microsurgery procedure was successfully conducted using the RAMS telerobotic system. The procedure demonstrated was the removal of a microscopic 0.015 inch diameter particle from a simulated eyeball.

Features added to the RAMS system to enable successful performance of the eye surgery demonstration were foot switch operated indexed motion, a surgical instrument mounted on the slave robot tip and a pivoting shared control algorithm to automatically compensate for pitch and yaw orientation of the surgical instrument while the operator controls the x-, y-, z- and roll motions of the instrument. Figure 10 shows the RAMS system as seen performing the simulated eye microsurgery procedure.




Figure 10: Performing the eye surgery demonstration.

In the next year, the RAMS system will be upgraded to implement force feedback to the master arm from force sensors mounted on the slave robot. In addition, experiments will be conducted to characterize the performance of the system as copmpared to direct manual manipulation in simulations of microsurgical tasks.

# References

- Charles, S. "Dexterity Enhancement for Surgery", in Computer Integrated Surgery: Technology and Clinical Applications, ed. R. H. Taylor, S. Lavalle, G. Burdea, R. Mosges, MIT Press, Cambridge, MA 1996.
- [2] Ousterhout, J. K. "Tcl and the Tk Toolkit", Addison Wesley, Reading, Mass. 1994.
- [3] Real-time Innovations, Inc., "Control Shell Programmer's Reference Manual Vol. 1", Sunnyvale, CA, 1995.
- [4] Real-time Innovations, Inc., "Control Shell Programmer's Reference Manual Vol. 2", Sunnyvale, CA, 1995.
- [5] Real-time Innovations, Inc., "NDDS Programmer's Reference Manual", Sunnyvale, CA, 1995.
- [6] Rodriguez, G., K. Kreutz, and A. Jain, "A Spatial Operator Algebra for Multibody System Dynamics," The Journal of the Astronautical Sciences, Vol. 40, No. 1, pp. 27-50, January-March 1992.
- [7] Rodriguez, G., "Kalman Filtering, Smoothing, and Recursive Robot Arm Forward and Inverse Dynamics," IEEE Transactions on Robotics and Automation, Vol. 3, pp. 624-639, Dec. 1987.
- [8] Rosheim, M. E., "Robot Wrist Actuators", John Wiley & Sons., New York, 1989.
- [9] Schenker, P., Das, H., and Ohm, T. "A new robot for high dexterity microsurgery" Proceedings of the First International Conference, CVRMed '95, Nice, France April, 1995. also in Computer Vision, Virtual Reality and Robotics in Medicine, Lecture Notes in Computer Science, Ed. Nicholas Ayache, Springer-Verlag, Berlin 1995.
- [10] Williams III, R. L., "Forward and Inverse Kinematics of Double Universal Joint Robot Wrists," Space Operations, Applications and Research (SOAR) Symposium, Albuquerque, NM, June 26-28, 1990.

#### Acknowledgment

This work was carried out at the Jet Propulsion Laboratory under contract with the National Aeronautics and Space Administration. The authors affiliated with JPL are in the Automation and Control Section, Jet Propulsion Laboratory, California Institute of Technology, 4800 Oak Grove Drive, Pasadena, CA 91109. Steve Charles, MD is the CEO of MicroDexterity Systems, Inc.